



Vitalograph (Ireland) Limited, Gort Road Business Park, Ennis, Co. Clare, Ireland
Telephone: 065 686 4100 Fax: 065 682 9289

Certificate of Cross-Infection Efficiency

This is to certify that Vitalograph BVF (Bacterial Viral Filters for Respiratory Function Testing) Model No. 2820 have the Bacterial Filter Efficiency (BFE) and Viral Filter Efficiency (VFE) stated below.

BFE (MRSA, TBC, etc)	> 99.999996%
VFE (Influenza, HN1, SARS-CoV-2/-1 etc.)	> 99.999711%

Frank Keane

CEO

19th March 2020