

DECLARATION OF CONFORMITY

DoC#: TF-1623, Rev 03.2

Legal Manufacturer:

Innovacon, Inc.

Legal Manufacturer's Address:

9975 Summers Ridge Road

San Diego, CA 92121 USA

Declares that the product Product Name and Model(s)

DMO-101	Innovacon™ MOP One Step Morphine Test Strip (Urine)		
	InstAlert™ MOP One Step Morphine Test Strip (Urine)		
	SureStep™ MOP One Step Morphine Test Strip (Urine)		
	SureStep ™ MOP 300 Urine Drug Test Strip		
DMO-102	Innovacon™ MOP One Step Morphine Test Device (Urine) InstAlert™ MOP One Step Morphine Test Device (Urine)		
	SureStep™ MOP One Step Morphine Test Device (Urine)		
	SureStep ™ MOP 300 Urine Drug Test Cassette		

as described above are in conformity with the requirements of the standards listed in Appendix 1, Applicable Standards and Guidelines.

Additional Information:

EC Representative's Name:

Medical Device Safety Service GmbH

EC Representative's Address:

Schiffgraben 41

30175 Hanover, Germany

Manufacturing Site:

ABON Biopharm (Hangzhou) Co., Ltd.

#198, 12th Street East

Hangzhou Economic and Technological Development Area

310018 Hangzhou, PR China

Management System:

MAN-003, Quality System Manual

Quality System Certificate No: SX 2236944-1

CV 2226044 4

Conformity Pathway:

Annex III

Classification:

Article 9, Section 1, Other IVD

EDMA Code:

12.70.09.08.00 - Opiates Rapid Test



This Declaration of Conformity is issued under the sole responsibility of Innovacon, Inc. I, the undersigned, hereby declare on behalf of the manufacturer, Innovacon, Inc., that the medical devices specified above conform with the Directive 98/79/EC on in vitro diagnostic medical devices and pertinent essential requirements.

Crystal Lee

Specialist II, Regulatory Affairs

San Diego, California, USA Date: July 18, 2022

Appendix 1 to DOC # TF-1623			
Applicable Standards and Guidelines			
Category :	Name	Number: Date Issued	
General	In Vitro Diagnostic Device Directive	98/79/EC: 27 Oct 1998	
	Medical Devices – Quality management systems –	EN ISO 13485:2016 +	
	Requirements for regulatory purposes	A11:2021	
Risk	Medical Devices – Application of risk management to	EN ISO 14971:2019+	
	medical devices	A11:2021	
	Medical Devices – Application of usability engineering to medical devices	EN 62366:2015	
	medical devices		
Labeling	Symbols for use in the labeling of medical devices	EN 15223-1:2021	
	In vitro diagnostic medical devices - Information supplied by	EN ISO 18113-1:2011	
	the manufacturer (labelling) – Part 1: Terms, definitions and		
	general requirements		
	In vitro diagnostic medical devices - Information supplied by	EN ISO 18113-2:2011	
	the manufacturer (labelling) – Part 2: In vitro diagnostic		
	reagents for professional use		
Performance	Performance evaluation of in vitro diagnostic medical	EN	
	devices	13612:2002/AC:2002	
	In vitro diagnostic medical devices – Evaluation of stability of	EN ISO 23640:2015	
	in vitro diagnostic reagents		