



**DECLARATION OF CONFORMITY**

<b>Legal Manufacturer:</b>	Becton, Dickinson and Company Limited, Donore Road, Drogheda, Co. Louth, Ireland.
<b>Authorised Representative:</b>	NA
<b>Manufacturing Site:</b>	Manufactured on behalf of Becton, Dickinson and Company, Limited by:  Becton Dickinson S.A. Crta. Mequinenza s/n. 22520 – Fraga HUESCA – SPAIN
<b>Products:</b>	BD Microlance™ 3 Needles (see attached list for the different types)
<b>Classification:</b>	Class IIa, Rule 6
<b>Conformity Assessment Route:</b>	Annex VII and Annex V
<b>GMDN:</b>	59230; sterile 59229; non-sterile

We herewith declare that for the above mentioned products, their product design meets the provisions of the European Medical Device Directive 93/42/EEC of 14 June 1993 and Irish Regulation; S.I. No. 252 of 1994 (as amended by Regulation 5 of S.I. 110/2009) concerning medical devices. All supporting documentation is retained at the premises of the manufacturer or subcontractor.

<b>Harmonized standards:</b>	EN 556-1:2001/AC:2006, EN 980:2008, EN 1041:2008/A1:2013, EN1707:1996, EN 20594-1:1993/AC:2009, EN ISO 10993-1:2009/AC:2010, EN ISO 10993-3:2014, EN ISO 10993-4:2009, EN ISO 10993-5:2009, EN ISO 10993-7:2008/AC2009, EN ISO 10993-11:2009, EN ISO 10993-12:2012, EN ISO 10993-13:2010, EN ISO 10993-15:2009, EN ISO 10993-17:2009; EN ISO 10993-18:2009; EN ISO 11135-1:2007, EN ISO 11138-2:2009, EN ISO 11737-1:2006/AC:2009, EN ISO 11737-2:2009, EN ISO 11607-1:2009, EN ISO 11607-2:2006, EN ISO 13485:2012, ISO13485:2003, EN 14155:2011, EN ISO 14971:2012,
<b>Non -Harmonised Standards</b>	ISO 14644-1:1999, ISO 10993-2:2006, ISO 10993-10:2009, ISO 9626: 1995, ISO 2859-1:1999, ISO 7864-1:1996; ISO15223-1:2012, ISO 594-1:1993, ISO 594-2:1998, ISO 6009; 1995/COR 2008
<b>Notified Body:</b>	NSAI 1 Swift Square, Northwood, Santry Dublin 9, Ireland Telephone: 353-1-807-3800 Number of Notified Body: 0050
<b>CE Certificate number:</b>	252.157
<b>Date of issue of original CE certificate:</b>	24 July 2012

Date: 23 November 2017

Date: 23 November 2017

Name Andrew Roche  
Function QA/RA  
Business Unit Becton, Dickinson and Company Limited

Cormac Reynolds  
Director  
Becton, Dickinson and Company Limited

**DECLARATION OF CONFORMITY**

**BD MICROLANCE NEEDLES**

**STERILE NEEDLES**

CATALOG No.	PRODUCT DESCRIPTION	
	(Imperial, inch)	(S.I.) mm
300094	22G x 2	0.7 x 50
300300	26G x 3/8	0.45 x 10
300400	25G x 1	0.5 x 25
300600	25G x 5/8	0.5 x 16
300635	27G x 1/2	0.4 x 13
300637	16G x 1 1/2	1.6 x 40
301155	21G x 2	0.8 x 50
301500	19G x 1 1/2	1.1 x 40
301700	19G x 1	1.1 x 25
301750	19G x 2	1.1 x 50
301900	18G x 2	1.2 x 50
302200	27G x 3/4	0.4 x 19
303800	26G x 1/2	0.45 x 13
304000	30G x 1/2	0.3x 13
302809*	30G x 1/2	0.3x 13
304100	24G x 1	0.55 x 25
304300	26G x 5/8	0.45 x 16
304622	18G x 1 1/2	1.2 x 40

**\*Pharmaceutical Quality**

## DECLARATION OF CONFORMITY

### BD MICROLANCE NEEDLES

#### NON-STERILE NEEDLES (BULK)

CATALOG No.	PRODUCT DESCRIPTION	
	(Imperial, inch)	(S.I.) mm
300206	25G x 5/8"	0.5 x 16
300284	21G x 5/8	0.8 x 16
300403	25G x 1	0.5 x 25
300801	23G x 1	0.6 x 25
300802	21G x 1 ½	0.8 x 40
300806	21Gx 1	0.8 x 25
300807	20Gx 1	0.9 x 25
300808	20G X 1 ½	0.9 x 40
300901	22G X 1 ¼	0.7 x 30
300932	30G X ½	0.3 x 13
301296	27G X ½	0.4 x 13
301530	19G X 1½	1.1 x 40
302424	14G X 1 ¼	2.1 x 30
303803	26G X ½	0.45 x 13
304623	26G X 3/8	0.45 x 10
304627	23G X 1	0.6 x 25
304630	21Gx2	0.8 x 50
304636	18G X 2	1.1 x 50
304637	27G X ¾	0.4 x 19
304641	24G X 1	0.55 x 25
304644	18G X 1½ SB TW	1.2 x 40
304728	22G X 1	0.7 x 25
304729	22G X 1½	0.7 x 40
305146	23G X 1 ¼	0.6 x 30
306030	25G x 5/8	0.5 x 16