

# EC Declaration of Conformity

We, **Kaz Europe Sàrl**, Place Chauderon 18, 1003 Lausanne, Switzerland declare under our sole responsibility that the product to which this declaration relates is in conformity with the following standard(s) or other normative document(s) and proves the conformity of the designated product with the provisions of the below Directive(s):

- **MDD - Council Directive 93/42/EEC of 14 June 1993 concerning medical devices**
- **RoHS - Directive 2011/65/EU of the European Parliament and of the Council of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment**

## Name

Braun ThermoScan 7 /ThermoScan 5 / Thermoscan 6 /  
Thermoscan 7+: IRT6520, IRT6030, IRT6020, IRT6515 and  
IRT6525 series

## Type or model

IRT6520	IRT6020NOEE
IRT6520MNLA	IRT6020MNLA
IRT6520WE	IRT6030
IRT6520BWE	IRT6515NOEE
IRT6520NOEE	IRT6515MNLA
IRT6520LA	
IRT6525	
IRT6525NOEE	
IRT6525MNLA	
IRT6525WE	
IRT6525KO	
IRT6525AP	
IRT6525AU	

IRT6520WEGP (IRT6520WE + a Toy thermometer)  
IRT6520NOEEGP (IRT6520EEE + a Toy thermometer)  
IRT6520MNLAP (IRT6520MNLA + a Toy thermometer)

Note: IRT6520WEGP, IRT6520NOEEGP, IRT6520MNLAP are sold with a Toy thermometer into the packaging. The toy is covered by his own EC Declaration of Conformity under the Toy Safety Directive with Lechner Ges.m.b.H, Osterreich as manufacturer.

## Standards Applied:

Standard Reference	Edition	Title
EN ISO 13485	2016	Medical devices — Quality management systems — Requirements for regulatory purposes
EN 60601-1	2006 +A1:2013	Medical electrical equipment - Part 1: General requirements for safety and essential performance.
EN 60601-1-2	2015	Medical electrical equipment – part 1-2: General requirements for basic safety and essential performance – Collateral standard: electromagnetic compatibility – Requirements and tests.
EN 60601-1-6	2010/A1:2013	Medical electrical equipment — Part 1-6: General requirements for basic safety and essential performance — Collateral Standard: Usability.
EN 60601-1-11	2015	Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance - Collateral standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment
EN ISO 14971	2019	Medical devices — Application of risk management to medical devices.
EN 62304	2006 A1:2008	Medical device software - Software life-cycle processes

# EC Declaration of Conformity

EN ISO 10993-1	2009/AC:2010	Biological evaluation of medical devices - Part 1: Evaluation and testing.
EN ISO 10993-5	2009	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity
ISO 10993-10	2010	Biological evaluation of medical devices – Part 10: Tests for irritation and skin sensitization
EN 62366-1	2015	Medical devices — Application of usability engineering to medical devices.
EN ISO 80601-2-56	2017	Medical electrical equipment – Part 2-56: Particular requirements for basic safety and essential performance of clinical
EN 12470-5	2000+A1:2009	Clinical thermometers - Part 2: Phase change type (dot matrix) thermometers
EN 1041	2008/A1:2013	Information supplied by the manufacturer with medical devices
EN ISO 15223-1	2016	Symbols to be used with medical device labels, labeling, and information to be supplied – Part 1 – General requirements
ASTM E1965-98	2016	Standard Specification for Infrared Thermometers for Intermittent Determination of Patient Temperature
EN ISO 14155	2011/AC:2011	Clinical investigation of medical devices for human subjects - Good clinical practice

The Technical Documentation is the responsibility of: Kaz Europe Sàrl, Place Chauderon 18, CH-1003 Lausanne, Switzerland

## Additional information:

<b>For Medical Device Directive 93/42/EC</b>	
<b>Regulatory class (MDD, Annex IX):</b>	class IIa (Annex IX rule 10)
<b>Conformity assessment procedure:</b>	Annex V
<b>GMDN</b>	17887
<b>UMDNS</b>	17-887
<b>Notified Body</b>	DQS Medizinprodukte GmbH August Schanz Str. 21 D-60433 Frankfurt, Germany Registration number: 0297
<b>EC Certificate</b>	381008 MR5
<b>EN ISO 13485 Certificate</b>	381008 MP2016

Authorized Representative in Europe:  
Address:

Obelis, S.A.  
Bd. Général Wahis, 53  
1030 Brussels, Belgium

Authorized Representative in Turkey:  
Address:

Sistem Çözüm Ortaklığı Satış Dağıtım Tic. Ltd. Şti.  
Ortaklar Cad. Bahçeler Sok.  
18 İş Merkezi K:3 D:5 Mecidiyeköy  
34394 İstanbul, Turkey  
+90 212 216 2950

Tel:

This declaration of conformity is valid until May 26, 2024.

Michael Burke



Lausanne

December 8, 2021

General Manager EMEA

Legally binding signature

Place

Date

Company Stamp:



Kaz Europe Sàrl  
Place Chauderon 18  
1003 Lausanne  
Switzerland  
Tel. +41 21 644 01 10  
Fax. +41 21 644 01 11



# EC-CERTIFICATE

(Production quality assurance)



This is to certify that the company

## Kaz Europe Sàrl

Place Chauderon 18  
1003 Lausanne  
Switzerland

has implemented and maintains a quality assurance system which applies to the manufacture and final controls of the products.

Through an audit, documented in a report, performed by DQS Medizinprodukte GmbH, it was verified that the management system fulfills the requirements of

## Annex V of Council Directive 93/42/EEC concerning medical devices

with respect to the following medical devices:

Devices for vital parameter monitoring according to annex

The manufacturer is subject to surveillance according to Annex V, Section 4. The CE marking with the Notified Body Identification Number (0297) may be affixed on the devices listed in the certificate. The certificate is in the case of class I(s) devices (I(s) = class I products placed on the market in sterile conditions) limited to the aspects of manufacture concerned with securing and maintaining sterile conditions. The certificate is in the case of class I(m) devices (I(m) = class I devices with a measuring function) limited to the aspects of manufacture concerned with the conformity of the products with the metrological requirements.

Certificate registration no.	381008 MR5
Certificate unique ID	170774273
Effective date	2021-04-21
Expiry date	2024-05-26
Frankfurt am Main	2021-04-21

## DQS Medizinprodukte GmbH

Sigrid Uhlemann  
Managing Director

Dr. Thomas Feldmann  
Head of Certification Body

August-Schanz-Straße 21, 60433 Frankfurt am Main,  
Tel. +49 (0) 69 95427-300, [medical.devices@dqs-med.de](mailto:medical.devices@dqs-med.de)

DQS Medizinprodukte GmbH is a Notified Body according to Council Directive 93/42/EEC concerning medical devices with the Identification Number 0297.



**Annex to certificate**  
**Certificate registration No.: 381008 MR5**  
**Certificate unique ID: 170774273**  
**Effective date: 2021-04-21**



**Kaz Europe Sàrl**

Place Chauderon 18  
1003 Lausanne  
Switzerland

<b>Device family</b>	<b>Device</b>	<b>Class</b>	
Devices for vital-parameter monitoring	Vicks/Wick IR Thermometer (Forehead) VFH100 and WFH100 series, BST200	Ila	
	Vicks No Touch Thermometer VNT200	Ila	
	Braun Blood Pressure Monitor (wrist), BBP2000 and BBP2200	Ila	
	Braun Blood Pressure Monitor (wrist) Type BPW4500	Ila	
	Braun Blood Pressure Monitor (upper arm) BP6000 series (BP6000, BP6100 and BP6200)	Ila	
	Braun ExactFit™ 3 / ExactFit™ 5 Blood Pressure Monitor BP6000 series (BUA6150WE, BUA6150CEME, BUA6350, BP6200PHEMEAV1)	Ila	
	Braun No Touch + Forehead Thermometer (also named as Braun Touchless + Forehead Thermometer) NTF3000 series NTF 3000 NTF3000WE NTF3000EE NTF3000AP NTF3000KO NTF3000AU NTF3000CN BNT400 BNT300	Ila	
	Braun IR Thermometer Type 3000, IRT3030	Ila	
	Braun Digital Thermometer Type 1000, PRT1000, PRT2000	Ila	
	Protection Cap for IRT thermometer Type LF20, PC20, LF40 (double pack)	Ila	
	Braun IR Thermometer Type 6000, IRT6020, IRT6520, IRT6030, IRT6515, IRT6525	Ila	
	Devices for vital-parameter monitoring	Braun Blood Pressure Monitor (upper arm) BUA5000 BUA5000EU BUA5000LA BUA5000LAD1 BUA7200	Ila
		Braun® BNA100 Nasal Aspirator	Ila